


2
Municipal Corporation Bhilai
NIT

77 एम.एल.डी. जलशोधन प्रयोगशाला में ई-कोलिफॉर्म बैक्टीरिया टेस्ट हेतु
आवश्यक उपकरण कय करने का कार्य ।

S.No.	Particulars	Qty	Rate
1	Laminar air flow 2x2x3 Construction-Outer body made of M.S. powder coated & amp; Inner working area made of S.S. HEPPA filter-Efficiency of heppa filter of 99.97 down on 0.3 microns and amp; pre filter down to 5 micron suitable blower fitted at backside of uniform flow. Side Arcylic panel Door fitted with acrylic door Supplied with 1 U.V. light & amp; 1 day light. Supplied with 1 analogue manometer to measure pressure flow supplied with gas cork.	1	
2	Filtration assembly with vacuum pump single	1	
3	Water bath 12 ho'es	1	
4	Autoclave 45 Lit	1	


Executive Engineer
Municipal Corporation
Bhilai(C.G.)


Assistant Engineer
Municipal Corporation
Bhilai(C.G.)


Sub Engineer
Municipal Corporation
Bhilai(C.G.)